



Certificate

Product Name

Surface Salinity Meter

Model

SSM-21P

Manufacturer

DKK-TOA CORPORATION

Test Report No.

KKEA9008

The Model SSM-21P manufactured by DKK-TOA Corporation was tested to evaluate equivalence for measuring the level of contamination of soluble salts on surfaces in comparison to the Bresle method as defined by ISO 8502-9 under NACE Standard SP0508-2010.

The measurement results obtained from the Model SSM-21P were deemed equivalent to those obtained by using the Bresle method. Hence use of the Model SSM-21P is equivalent to ISO 8502-9.

Date: 21st Feb. 2011

Certified by Nobuaki NAKASHIMA
Certified Environmental Measurer

Sumitomo Heavy Industries Environment Co., ltd. Development and Analysis Center

19, Natsushima-cho, Yokosuka-shi, Kanagawa, 237-8555, Japan Tel.046-869-2450 Fax.046-869-1793